

Examination of methods to separate overlapping metabolites at 7T

Article

Published Version

Creative Commons: Attribution-Noncommercial-No Derivative Works 4.0

Open Access

Bell, T. K. ORCID: <https://orcid.org/0000-0002-9591-706X>,
Goerzen, D. ORCID: <https://orcid.org/0000-0001-9645-3503>,
Near, J. ORCID: <https://orcid.org/0000-0003-3516-936X> and
Harris, A. D. ORCID: <https://orcid.org/0000-0003-4731-7075>
(2024) Examination of methods to separate overlapping
metabolites at 7T. *Magnetic Resonance in Medicine*, 93 (2).
pp. 470-480. ISSN 0740-3194 doi: 10.1002/mrm.30293
Available at <https://centaur.reading.ac.uk/125526/>

It is advisable to refer to the publisher's version if you intend to cite from the work. See [Guidance on citing](#).

To link to this article DOI: <http://dx.doi.org/10.1002/mrm.30293>

Publisher: Wiley

All outputs in CentAUR are protected by Intellectual Property Rights law,
including copyright law. Copyright and IPR is retained by the creators or other
copyright holders. Terms and conditions for use of this material are defined in
the [End User Agreement](#).

www.reading.ac.uk/centaur

CentAUR

Central Archive at the University of Reading

Reading's research outputs online

Examination of methods to separate overlapping metabolites at 7T

Tiffany K. Bell^{1,2,3}   | Dana Goerzen⁴  | Jamie Near^{5,6}  | Ashley D. Harris^{1,2,3} 

¹Department of Radiology, University of Calgary, Calgary, Alberta, Canada

²Hotchkiss Brain Institute, University of Calgary, Calgary, Alberta, Canada

³Alberta Children's Hospital Research Institute, University of Calgary, Calgary, Alberta, Canada

⁴Weill Cornell Medicine, Cornell University, New York, New York, USA

⁵Physical Studies Research Platform, Sunnybrook Research Institute, Toronto, Ontario, Canada

⁶Department of Medical Biophysics, University of Toronto, Toronto, Ontario, Canada

Correspondence

Tiffany K. Bell, Alberta Children's Hospital, 28 Oki Drive, Calgary, AB T3B 6A8, Canada.

Email: tiffany.bell@ucalgary.ca

Funding information

Natural Sciences and Engineering Research Council of Canada, Grant/Award Numbers: RGPIN-2017-03875, RGPIN-2020-05917; New Frontiers Research Fund of Canada, Grant/Award Number: NFRFT 2022-00327; Canadian Foundation for Innovation, John R. Evans Leaders Fund; Canadian Institutes of Health Research, Grant/Award Number: PJT-183715; Canada Research Chairs, Magnetic Resonance Spectroscopy in Brain Injury, Grant/Award Number: PJ-T175085

Abstract

Purpose: Neurochemicals of interest quantified by MRS are often composites of overlapping signals. At higher field strengths (i.e., 7T), there is better separation of these signals. As the availability of higher field strengths is increasing, it is important to re-evaluate the separability of overlapping metabolite signals.

Methods: This study compares the ability of stimulated echo acquisition mode (STEAM-8; TE = 8 ms), short-TE semi-LASER (sLASER-34; TE = 34 ms), and long-TE semi-LASER (sLASER-105; TE = 105 ms) acquisitions to separate the commonly acquired neurochemicals at 7T (Glx, consisting of glutamate and glutamine; total N-acetyl aspartate, consisting of N-acetyl aspartate and N-acetylaspartylglutamate; total creatine, consisting of creatine and phosphocreatine; and total choline, consisting of choline, phosphocholine, and glycerocephocholine).

Results: sLASER-34 produced the lowest fit errors for most neurochemicals; however, STEAM-8 had better within-subject reproducibility and required fewer subjects to detect a change between groups. However, this is dependent on the neurochemical of interest.

Conclusion: We recommend short-TE STEAM for separation of most standard neurochemicals at 7T over short-TE or long-TE sLASER.

KEY WORDS

7 T, brain, high field, magnetic resonance spectroscopy, semi-LASER, STEAM

1 | INTRODUCTION

Proton MRS is commonly used to quantify chemicals of interest in the human brain. The typically quantified neurochemicals are Glx, total N-acetyl aspartate (tNAA), total creatine (tCr), and total choline (tCho). However, these are composites of signals from multiple neurochemicals that overlap. For example, the Glx signal consists of the overlapping signals' glutamate (Glu) and glutamine (Gln).

Higher field strength increases SNR and spectral resolution, thus potentially allowing better separation of overlapping signals. As the availability of higher field strengths (i.e., 7T) increases, it is important to re-evaluate the separability of overlapping metabolite signals. The most commonly used acquisition protocol at 7T is stimulated echo acquisition mode (STEAM). STEAM allows for ultrashort TEs, which minimize signal modulation due to *j*-coupling and T_2 relaxation, desirable aspects for separating overlapping signals.¹ For example, Zhong and Ernst showed improvement of spectral quality and reduced signal loss for the coupled resonances Glu and Gln, as well as myo-inositol at $TE = 8$ ms compared with $TE = 30$ ms.² Dou et al. also showed better separation of Glu and Gln signal using a short-echo STEAM sequence ($TE = 20$ ms) compared with a long-echo STEAM sequence ($TE = 74$ ms) at 7T.³ However, STEAM results in a lower SNR (compared to spin echo acquisitions), because only half the signal can be obtained.

Although semi-LASER (sLASER) is unable to achieve the very short TEs of STEAM due to the need for longer adiabatic pulses, sLASER is able to achieve higher SNR. sLASER also results in reduced chemical shift displacement and better resilience to B_1 inhomogeneity, both important aspects when working at higher field strengths.⁴ Furthermore, the paired adiabatic pulses in sLASER suppress *j*-evolution and prolong T_2 relaxation times, similar to the effects of short-TE sequences.⁵ Indeed, Marsman et al. reported that sLASER ($TE = 28$ ms) produces more reproducible measures of Glu than STEAM ($TE = 7.5$ ms) at 7T.⁶ However, Okada et al. found no difference in repeatability when comparing sLASER ($TE = 32$ ms) with STEAM ($TE = 5$ ms). Additionally, although sLASER did have a higher N-acetyl aspartate (NAA) SNR, they found that STEAM produced lower Cramér-Rao lower bounds (CRLBs) for several metabolites.⁷ Therefore, the optimal sequence and TE may depend on the metabolite of interest.

Although short TEs are typically preferred to prevent signal loss, longer TEs may sometimes be advantageous for detecting certain metabolites, particularly if optimal pulse timings are chosen to yield favorable peak detection.^{8,9} Shorter TEs will also result in more macromolecule signal, which can complicate metabolite quantification. Furthermore, to achieve short TEs when using sLASER at

7T, special hardware modifications are sometimes needed, such as a dual transmit option,⁶ which may not be available. Wong et al. looked at the signal evolution of Glu over a range of TEs from 45 ms to 225 ms and found the optimal TE to be 105 ms.¹ However, this was based on Glu only, and TEs below 45 ms were not investigated. Indeed, Najac et al. showed the highest Glu signal to be at $TE = 38$ ms, followed by $TE = 100$ ms¹⁰; however, this trend of increasing signal at higher TEs was not seen for other metabolites.

The aim of this study, therefore, was to compare the ability of STEAM, short-TE sLASER, and long-TE sLASER acquisitions to separate the commonly acquired neurochemicals at 7T.

2 | METHODS

2.1 | Participants

Fourteen healthy participants between the ages of 18 and 40 years were recruited. Data were collected with approval from the local research ethics board, and written informed consent was obtained from all participants.

2.2 | Data acquisition

Data were collected using a 7T Siemens Terra with a 32-channel head coil. Each session began with a T_1 -weighted image (MPRAGE, $TE/TR = 1.84/4300$ ms, isotropic resolution = 1 mm³) for voxel placement and segmentation. The 2.5 × 2.5 × 2.5 cm³ voxels were placed in the parietal cortex centered on the midline (Figure 1).

Data were collected using three sequences: STEAM-8 ($TE/TR/TM = 8/6000/40$ ms), sLASER-34 ($TE/TR = 34/5000$ ms), and sLASER-105 ($TE/TR = 105/5000$ ms), all with 8224 spectral points and outer-volume suppression pulses applied in all cases. For each sequence, 64 water-suppressed averages were acquired using VAPOR water suppression, and 8 water-unsuppressed averages were acquired. Sequence order was counterbalanced across subjects. The STEAM sequence¹¹ was developed by Edward J. Auerbach and Małgorzata Marjańska, and the sLASER sequence^{12,13} was developed by Gülin Öz and Dinesh Deelchand and provided by the University of Minnesota under a C2P agreement.

2.3 | Data processing

MRS data were preprocessed using FID-A¹⁴ with the following steps: coil combination, removal of motion-corrupted averages, frequency drift correction,




FIGURE 1 (A) Example of voxel location in the parietal cortex. Mean (black) and SD (shaded gray) of spectra acquired using (B) STEAM-8 (stimulated echo acquisition mode, TE = 8 ms), (C) sLASER-34 (semi-LASER, TE = 34 ms), and (D) sLASER-105 (semi-LASER, TE = 105 ms).

and zero-order phase correction. LCModel¹⁵ was used to apply eddy current correction and quantification relative to water. The ATT_{H2O} parameter was set to 1, to allow for correction of T₂ decay of water at each TE external to LCModel.

Basis sets for quantification were simulated using FID-A based on exact timings and RF pulses for each sequence. Basis sets included the following metabolites: alanine, aspartate, choline (Cho), glycerophosphocholine (GPC), phosphocholine (PC), creatine (Cr), phosphocreatine (PCr), γ -aminobutyric acid, glutamate (Glu), glutamine (Gln), lactate, myo-inositol, NAA, N-acetylaspartylglutamate (NAAG), scyllo-inositol, glutathione, glucose, taurine, glycine, and phosphatidylethanolamine (basis sets, including simulated macromolecules, are available at <https://github.com/HarrisBrainLab/BasisSets>).

Macromolecule basis sets were simulated using information obtained by parameterizing freely available 7T metabolite-suppressed data from 6 healthy individuals (<https://zenodo.org/records/3906754#.XvOWVvJ7n7e>).¹⁶ First, all six metabolite-nulled spectra were combined into a single average macromolecule spectrum.

Nine macromolecule resonances were then identified in the average macromolecule spectrum, and each macromolecule resonance was fit to a Gaussian lineshape using in-house custom *MATLAB* scripts to estimate the frequency, linewidth, and relative amplitude. For each identified macromolecule resonance, a single-proton FID-A spin system was created, with the “scaleFactor” parameter adjusted to account for the number of protons associated with each macromolecule resonance (see Table 1; FID-A spin systems are available at <https://github.com/HarrisBrainLab/MMSim>). Macromolecules were then simulated with the appropriate simulation function in FID-A based on the parameters in Table 1, using exact timings and RF pulses as described previously. The simulated macromolecule linewidth was 6 Hz smaller than the linewidths of the parameterized spectra to account for the line broadening typically applied by LCModel, to match the narrow, simulated metabolite basis functions to the typically broader linewidths observed in *in vivo* data. The relative concentration of each macromolecule was constrained using the values in Table 1. Macromolecule basis sets were included for quantification of STEAM-8 and sLASER-34 data. At 105 ms, the macromolecule

TABLE 1 Values used for simulation of macromolecules.

Macromolecule	Frequency (ppm)	No. protons ^a	Simulated linewidth (Hz)	LCModel constraint ^b
M _{0.95}	0.95	3	32	n/a
M _{1.24}	1.235	3	23	0.41 ± 0.16
M _{1.43}	1.427	3	26	0.98 ± 0.67
M _{1.71}	1.705	2	31	1.185 ± 1.10
M _{2.07}	2.07	2	38	2.53 ± 0.81
M _{2.30}	2.295	2	32	2.00 ± 0.17
M _{3.03}	3.03	2	26	1.30 ± 0.32
M _{3.24}	3.24	2	48	0.10 ± 0.22
M _{3.98}	3.98	2	109	2.33 ± 0.72

Abbreviation: n/a, not applicable.

^aValues obtained from Cudalbu et al.²²

^bConcentration relative to M_{0.95}.

signal had decayed to approximately 0; thus, there is risk of overfitting if including macromolecule basis sets. Therefore, sLASER-105 was quantified with and without the macromolecule basis set. All results were consistent; thus, we report sLASER-105 quantification without macromolecule basis sets.

T₁-weighted images were segmented using the CoReg-StandAlone function from the Gannet toolbox.¹⁷ Tissue correction was performed according to Gasparovic et al.,¹⁸ including correcting for the different TEs and T₁ and T₂ values for both water and metabolites (see Table S3 for T₁ and T₂ values used). SNR was estimated using the op_getSNR function in FID-A, which measures the height of the NAA resonance between 1.8 and 2.2 ppm and normalizes this to the SD of the signal in a peak free region (manually set to between -1 and -2 ppm).

Consistent with the field, we used within-subject variation as our metric of measurement precision (see Supporting Information for simulations comparing the effects of measurement precision on within-subject and between-subject variability). To assess repeatability, the water-suppressed scans were split into two equal blocks of 32 averages, and the water-unsuppressed scans were split into two equal blocks of four averages, then processed as described previously. These split scans were used for calculation of the within-subject coefficient of variation (CV) only; all other metrics were calculated using the full 64-average scans.

2.4 | Statistical analysis

Statistical analyses were conducted using R (version 4.1.2; R Core Team [2020] <https://www.R-project.org/>).¹⁹ The analysis focused on the following combinations of metabolites: Glx, Glu + Gln; tNAA, NAA + NAAG;

tCr, Cr + PCr; and tCho, Cho + PC + GPC. Due to the increasing interest in measuring glutathione (GSH) and γ-aminobutyric acid (GABA) at high fields, results for these metabolites are reported in the Supporting Information.

Quality metrics (linewidth, SNR), metabolite levels, and CRLBs obtained from the full 64-average scan were statistically compared using repeated-measures analyses of variance with the “aov” function in the “stats” package. Pairwise comparisons were conducted using Bonferroni correction. CRLB was included as a measure of fit precision, as poor fitting would be more likely to result in signal being incorrectly assigned to overlapping metabolites, particularly for lower concentration metabolites.

Repeatability was assessed using the split 32-average scans. The within-subject CV was calculated as the within-subject variance divided by the mean, using the RMS approach.²⁰ The 95% confidence intervals were calculated for the CV values using bootstrap resampling with replacement (1000 resamples). Additionally, the pairwise correlation coefficient from LCModel was used to assess the relationship between overlapping metabolites (e.g., the correlation between NAA and NAAG). A correlation closer to zero implies better separation of overlapping signal, whereas positive or negative correlations with a larger magnitude indicate poor separation,³ and it is recommended in the LCModel manual that signal has not been adequately separated if the correlation between a pair is consistently less than -0.3.

2.5 | Sample-size calculations

Calculations were performed for each metabolite to estimate the number of participants per group that would be

needed to detect a 10% change between two groups, based on the formula described by Wong et al.¹

$$n = (u + v)^2 \times \frac{(2 \times \sigma^2)}{(\Delta \times \mu)^2} \quad (1)$$

where $\mu = 1.28\%$ for 90% power; $v = 1.96\%$ for 95% confidence level; $\Delta = 0.1$ for a 10% change; $\sigma =$ the SD of each metabolite across all subjects; and $\mu =$ the mean of each metabolite across all subjects.

3 | RESULTS

3.1 | Quality metrics

There was a significant effect of sequence on NAA linewidth ($p < 0.001$). Pairwise comparisons showed a significant difference for all comparisons (all $p < 0.001$), with sLASER-105 having the lowest NAA linewidth. There was also a significant effect of sequence on H₂O linewidth ($p < 0.01$). Pairwise comparisons showed that sLASER-105 had significantly lower H₂O linewidth than STEAM-8 ($p = 0.02$) and sLASER-34 ($p = 0.02$), but there was no significant difference between STEAM-8 and sLASER-34 ($p = 0.98$). There was also a significant effect of sequence on SNR ($p < 0.001$). Pairwise comparisons showed that sLASER-34 had significantly higher SNR than STEAM-8 ($p < 0.001$) and sLASER-105 ($p < 0.001$). There was no significant difference in SNR between STEAM-8 and sLASER-105 ($p = 0.43$, Table 2).

3.2 | Glutamate

Glu: There was a significant effect of sequence on metabolite level ($F[2,38] = 178.8$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for all three comparisons (all $p < 0.001$) (Figure 2A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 76.4$; $p < 0.001$). Glu measured using sLASER-105 had significantly lower CRLBs compared

TABLE 2 Mean and SDs of quality metrics for each sequence.

	NAA linewidth (Hz)	H ₂ O linewidth (Hz)	SNR
STEAM-8	13.21 \pm 1.65	11.73 \pm 1.82	254.47 \pm 48.23
sLASER-34	10.93 \pm 1.42	11.62 \pm 1.24	402.99 \pm 42.52
sLASER-105	8.54 \pm 1.05	9.89 \pm 1.87	276.53 \pm 44.90
<i>p</i>	< 0.001	0.009	< 0.001

Abbreviations: NAA, N-acetyl aspartate; sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms.

with Glu measured using the other sequences (Figure 2B). STEAM-8 produced the lowest CV (Figure 2C) and required the least number of participants to detect a 10% difference (Table 3).

Gln: There was a significant effect of sequence on metabolite level ($F[2,38] = 77.56$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for all three comparisons (all $p < 0.001$) (Figure 2A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 39.13$; $p < 0.001$). STEAM-8 produced significantly lower CRLBs compared with the other sequences (Figure 2B). STEAM-8 produced the lowest CV (Figure 2C) and required the least number of participants to detect a 10% difference (Table 3).

Glx: There was a significant effect of sequence on metabolite level ($F[2,38] = 117.2$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and STEAM-8 versus sLASER-105 ($p < 0.001$) (Figure 2A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 342.5$; $p < 0.001$). STEAM-8 produced the lowest CRLBs; however, this was not significantly different from sLASER-34 (Figure 2B). STEAM-8 produced the lowest CV (Figure 2C) and required the least number of participants to detect a 10% difference (Table 3).

The mean (\pm SD) of the correlation coefficients produced by LCModel for Glu and Gln for each sequence are as follows: STEAM-8 = 0.02 ± 0.02 , sLASER-34 = -0.01 ± 0.05 , sLASER-105 = -0.01 ± 0.32 .

3.3 | N-acetyl-aspartate

NAA: There was a significant effect of sequence on metabolite level ($F[2,38] = 36.22$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for sLASER-34 versus sLASER-105 ($p < 0.001$) and STEAM-8 versus sLASER-34 ($p < 0.001$) (Figure 3A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 12.1$; $p < 0.001$). Pairwise comparisons showed that sLASER-34 produced significantly lower CRLBs than the other two sequences ($p < 0.01$) (Figure 3B). STEAM-8 produced the lowest CV (Figure 3C). sLASER-34 required the least number of participants to detect a 10% difference (Table 3).

NAAG: There was a significant effect of sequence on metabolite level ($F[2,38] = 55.91$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and STEAM-8 versus sLASER-105 ($p < 0.001$) (Figure 3A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 5.0$; $p = 0.01$). STEAM-8 produced the lowest CRLBs, but these were not significantly different from




FIGURE 2 (A) Mean and spread of glutamate (Glu), glutamine (Gln), and Glu + Gln (Glx) data for each sequence. (B) Mean Cramér-Rao lower bound (CRLB) values for Glu, Gln, and Glx, error bars represent standard deviation. (C) Coefficient of variation (CV) values for Glu, Gln, and Glx, error bars represent 95% confidence intervals, calculated using bootstrap resampling. $*p < 0.05$, $**p < 0.01$, $***p < 0.001$. sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms.

TABLE 3 Estimated number of participants per group needed to observe a 10% difference between groups for each sequence and each metabolite.

	STEAM-8	sLASER-34	sLASER-105
Glu	6	13	21
Gln	47	137	153
Glx	6	15	23
NAA	6	7	25
NAAG	21	76	174
tNAA	6	9	27
Cr	6	4	40
PCr	15	60	38
tCr	1	9	16
Cho	19	3342	901
GPC	7	27	162
tCho	7	22	47

Abbreviations: Cr, creatine; Glu, glutamate; Gln, glutamine; Glx, Glu + Gln; GPC, glycerophosphocholine; NAA, N-acetyl aspartate; NAAG, N-acetyl aspartylglutamate; PCr, phosphocreatine; sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms; tCho, total choline; tCr, total creatine; tNAA, total N-acetyl aspartate.

sLASER-34 (Figure 3B). STEAM-8 produced the lowest CV (Figure 3C) and required the least number of participants to detect a 10% difference (Table 3).

tNAA: There was a significant effect of sequence on metabolite level ($F[2,38] = 40.43$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and sLASER-34 versus sLASER-105 ($p < 0.001$) (Figure 3A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 11.28$; $p < 0.001$). sLASER-105 produced the lowest CRLBs; however, these were not significantly different than sLASER-34 (Figure 3B). STEAM-8 had the lowest CV (Figure 3C). sLASER-34 required the least number of participants to detect a 10% difference (Table 3).

The mean (\pm SD) of the correlation coefficients produced by LCModel for NAA and NAAG for each sequence are as follows: STEAM-8 = -0.04 ± 0.07 , sLASER-34 = -0.16 ± 0.10 , and sLASER-105 = -0.61 ± 0.30 .

3.4 | Creatine

Cr: There was a significant effect of sequence on metabolite level ($F[2,38] = 86.71$; $p < 0.001$), and pairwise




FIGURE 3 (A) Mean and spread of N-acetyl aspartate (NAA), N-acetylaspartylglutamate (NAAG), and total N-acetyl aspartate (tNAA; NAA + NAAG) data for each sequence. (B) Mean Cramér-Rao lower bound (CRLB) values for NAA, NAAG and tNAA, error bars represent standard deviation. (C) Coefficient of variation (CV) values for NAA, NAAG and tNAA, error bars represent 95% confidence intervals calculated using bootstrap resampling. $*p < 0.05$, $**p < 0.01$, $***p < 0.001$. sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms.

comparisons showed a significant difference in metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and STEAM-8 versus sLASER-105 ($p < 0.001$) (Figure 4A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 15.7$; $p < 0.001$). sLASER-34 produced significantly lower CRLBs compared with the other two sequences (Figure 4B). STEAM-8 produced the lowest CV (Figure 4C). sLASER-34 required the least number of participants to detect a 10% difference (Table 3).

PCr: There was a significant effect of sequence on metabolite level ($F[2,38] = 26.88$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and STEAM-8 versus sLASER-105 ($p < 0.001$) (Figure 4A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 47.4$; $p < 0.001$). sLASER-34 produced significantly lower CRLBs compared with the other two sequences (Figure 4B). STEAM-8 produced the lowest CV (Figure 4C). STEAM-8 required the least number of participants to detect a 10% difference (Table 3).

tCr: There was a significant effect of sequence on metabolite level ($F[2,38] = 133.8$; $p < 0.001$), and pairwise comparisons showed a significant difference in

metabolite levels for STEAM-8 versus sLASER-34 ($p < 0.001$) and STEAM-8 versus sLASER-105 ($p < 0.001$) (Figure 4A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 178.9$; $p < 0.001$). sLASER-34 and sLASER-105 produced equally low CRLBs (Figure 4B). sLASER-105 produced the lowest CV (Figure 4C). STEAM-8 required the least number of participants to detect a 10% difference (Table 3).

The mean (\pm SD) of the correlation coefficients produced by LCModel for Cr and PCr for each sequence are as follows: STEAM-8 = -0.68 ± 0.02 , sLASER-34 = -0.71 ± 0.03 , and sLASER-105 = -0.88 ± 0.02 .

3.5 | Choline

Cho: There was a significant effect of sequence on metabolite level ($F[2,38] = 24.0$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for all three comparisons (all $p < 0.01$) (Figure 5A). There was also a significant effect of sequence on CRLBs ($F[2,32] = 17.4$; $p < 0.001$). STEAM-8 produced the




FIGURE 4 (A) Mean and spread of creatine (Cr), phosphocreatine (PCr), and total creatine (tCr; Cr + PCr) data for each sequence. (B) Mean Cramér-Rao lower bound (CRLB) values for Cr, PCr, and tCr, error bars represent SD. (C) Coefficient of variation (CV) values for Cr, PCr, and tCr, error bars represent 95% confidence intervals calculated using bootstrap resampling. * $p < 0.05$, ** $p < 0.01$, *** $p < 0.001$. sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms.

lowest CRLB, but this was not significantly lower than sLASER-105 (Figure 5B). STEAM-8 produced the lowest CV (Figure 5C) and required the least number of participants to detect a 10% difference (Table 3).

PC: PC was not quantified well using any of the three sequences, and therefore will not be included for comparison.

GPC: There was a significant effect of sequence on metabolite level ($F[2,38] = 112.0$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for all three comparisons (all $p < 0.01$) (Figure 5A). There was also a significant effect of sequence on CRLBs ($F[2,38] = 34.2$; $p < 0.001$). STEAM-8 produced the lowest CRLBs, but this was not significantly lower than sLASER-34 (Figure 5B). STEAM-8 produced the lowest CV (Figure 5C) and required the least number of participants to detect a 10% difference (Table 3).

tCho: There was a significant effect of sequence on metabolite level ($F[2,38] = 95.1$; $p < 0.001$), and pairwise comparisons showed a significant difference in metabolite levels for all three comparisons (all $p < 0.001$) (Figure 5A). There was no significant effect of sequence on CRLBs ($F[2,38] = 1.6$; $p = 0.22$) (Figure 5B). STEAM-8 produced

the lowest CV (Figure 5C) and required the least number of participants to detect a 10% difference (Table 3).

Due to PC not being quantified, only the correlation between Cho and GPC was assessed. The mean (\pm SD) of the correlation coefficients produced by LCModel for Cho and GPC for each sequence are as follows: STEAM-8 = 0.03 ± 0.05 , sLASER-34 = -0.02 ± 0.09 , and sLASER-105 = -0.05 ± 0.27 .

4 | DISCUSSION

Several studies have assessed the ability of STEAM and sLASER sequences to separate Glu and Gln signal at 7T. Here we expand on previous work to include the direct comparison of three sequences (STEAM-8, sLASER-34, and sLASER-105) to separate overlapping signal of the four most commonly quantified metabolites.

With the exception of creatine-containing metabolites, there was little correlation between most overlapping metabolites, suggesting that all sequences are generally able to resolve the metabolite signals (i.e., a high correlation between two metabolites would suggest limited or




FIGURE 5 (A) Mean and spread of choline (Cho), glycerophosphocholine (GPC) and total choline (tCho; Cho + phosphocholine (PC) + GPC) data for each sequence. (B) Mean Cramér-Rao lower bound (CRLB) values for Cho, GPC, and tCho, error bars represent SD. Mean Cho CRLB for sLASER-34 not shown as > 500 (C) Coefficient of variation (CV) values for Cho, GPC and tCho, error bars represent 95% confidence intervals calculated using bootstrap resampling. * $p < 0.05$, ** $p < 0.01$, *** $p < 0.001$. sLASER-34, semi-LASER, TE = 34 ms; sLASER-105, semi-LASER, TE = 105 ms; STEAM-8, stimulated echo acquisition mode, TE = 8 ms.

inconsistent success at signal separation). The magnitude of the correlation between Cr and PCr was greater than -0.3 for all sequences, indicating they are not well separated with any sequence, despite the low fit error, and are best reported as a composite measure. NAA and NAAG measured with sLASER-105 were also strongly correlated (-0.61 ± 0.30) and therefore should also be reported as a composite measure when using this sequence.

Based on fit error (CRLB), generally sLASER-34 appears to perform best at quantifying the individual neurochemicals as well as quantifying the summed signals. However, this is not true for choline-containing neurochemicals, as Cho and GPC quantified with STEAM-8 produced the lowest fit errors. STEAM-8 had better repeatability compared with the other sequences, and, generally, fewer participants were needed to detect a 10% change when using STEAM-8, although again this was dependent on the neurochemical.

Interestingly, metabolites quantified using STEAM-8 had less variation (lower CVs) than when quantified using the other two sequences. This is in contrast to Marsman et al., who found higher CVs with STEAM

(TE = 7.7 ms) compared to sLASER (TE = 28 ms).⁶ However, Marsman et al. compared variation across two different scan sessions rather than variation over the same session as done here. This implies that STEAM may produce more consistent results across a session but less consistent results across time. Because sLASER has a reduced chemical shift displacement and better resilience to B_1 inhomogeneity,⁴ it is likely less sensitive to repositioning effects.

STEAM-8 often produced higher values compared with the other two sequences, likely due to less signal loss from the small TE. There is also evidence that apparent T_2 values of metabolites may vary across sequences,²¹ which may also contribute to concentration differences, although this is not expected to largely affect the values. STEAM-8 generally also required fewer participants to detect a 10% change. This likely reflects a smaller spread of data when acquired using this sequence. As there is no ground truth when working with *in vivo* data, it is not possible to determine whether this reflects a more consistent measurement, or whether STEAM-8 is not picking up on smaller differences between participants.

Wong et al. recommend sLASER-105 for optimal Glu measurement; however, the shortest TE they included in their comparison was 45 ms. Although they observed the greatest *in vivo* glutamate signal at TE = 45 ms, they found more variability and larger CRLBs at this TE and suggest this to be due to variations in macromolecules.¹ We show lower variability and CRLBs using the shorter TEs, which will have the greatest macromolecule contamination. The inclusion of a simulated macromolecule baseline may have aided in controlling for this.

A limitation of this study is that only one area of the brain was assessed. Data were acquired from the parietal cortex, which generally produces good quality data. Areas of the brain where data are often of poorer quality (such as frontal areas or deep brain structures) may produce different results. Additionally, simulated macromolecule models were used to account for macromolecule signal, as opposed to directly measuring macromolecule signal *in vivo*. This method was chosen to reduce participant burden in the scanner and follows recommendations for accounting for macromolecule signal in metabolite spectra.²² However, the metabolite values (and subsequently the correlation and CRLB values) may be different when using a measured macromolecule signal, partly due to the increased degrees of freedom when modeling individually simulated macromolecules. This was partly mitigated in our analysis by using fixed proportions for each macromolecule, reducing the degrees of freedom.

5 | CONCLUSIONS

Overall, we show data acquired at 7T with sLASER-34 had lower fit errors, but data acquired with STEAM-8 had better within-session repeatability and required fewer subjects to detect a change between groups. However, this is dependent on the neurochemical of interest. We therefore recommend short-TE STEAM for separation of most standard neurochemicals over short-TE or long-TE sLASER.

ACKNOWLEDGMENTS

This work was undertaken thanks in part to funding from the Natural Sciences and Engineering Research Council of Canada (NSERC, RGPIN-2020-05917, RGPIN-2017-03875), a Canadian Foundation for Innovation John R Evans Leaders Fund (CFI-JELF), and support by the Hotchkiss Brain Institute and the Alberta Children's Hospital Research Institute, University of Calgary. T.K.B is supported by One Child Every Child and Cumming School of Medicine Postdoctoral Fellowships. J.N. holds funding from the Canadian Institutes of Health Research (CIHR, PJT-183715) and the New Frontiers Research Fund of Canada (NFRF, NFRFT 2022-00327).

A.D.H. holds a Canada Research Chair in Magnetic Resonance Spectroscopy in Brain Injury and holds funding from the CIHR (PJ-T175085) and The Arthritis Society.

DATA AVAILABILITY STATEMENT

Basis sets, including simulated macromolecules, are available at <https://github.com/HarrisBrainLab/BasisSets>. Tools to simulate macromolecules are available at <https://github.com/HarrisBrainLab/MMSim>. Anonymized data may be made available to qualified investigators upon request.

ORCID

Tiffany K. Bell  <https://orcid.org/0000-0002-9591-706X>

Dana Goerzen  <https://orcid.org/0000-0001-9645-3503>

Jamie Near  <https://orcid.org/0000-0003-3516-936X>

Ashley D. Harris  <https://orcid.org/0000-0003-4731-7075>

TWITTER

Tiffany K. Bell  [tiffanybello](https://twitter.com/tiffanybello)

REFERENCES

1. Wong D, Schranz AL, Bartha R. Optimized *in vivo* brain glutamate measurement using long-echo-time semi-LASER at 7 T. *NMR Biomed.* 2018;31:1-13. doi:10.1002/nbm.4002
2. Zhong K, Ernst T. Localized *in vivo* human 1H MRS at very short echo times. *Magn Reson Med.* 2004;52:898-901. doi:10.1002/mrm.20201
3. Dou W, Kaufmann J, Li M, Zhong K, Walter M, Speck O. The separation of Gln and Glu in STEAM: a comparison study using short and long TEs/TMs at 3 and 7 T. *Magn Reson Mater Phys Biol Med.* 2015;28:395-405. doi:10.1007/s10334-014-0479-7
4. Wilson M, Andronesi O, Barker PB, et al. Methodological consensus on clinical proton MRS of the brain: review and recommendations. *Magn Reson Med.* 2019;82:527-550. doi:10.1002/mrm.27742
5. Deelchand DK, Kantarci K, Öz G. Improved localization, spectral quality, and repeatability with advanced MRS methodology in the clinical setting. *Magn Reson Med.* 2018;79:1241-1250. doi:10.1002/mrm.26788
6. Marsman A, Boer VO, Luijten PR, et al. Detection of glutamate alterations in the human brain using 1H-MRS: comparison of STEAM and sLASER at 7 T. *Front Psych.* 2017;8:60. doi:10.3389/fpsyg.2017.00060
7. Okada T, Kurabayashi H, Kaiser LG, et al. Repeatability of proton magnetic resonance spectroscopy of the brain at 7 T: effect of scan time on semi-localized by adiabatic selective refocusing and short-echo time stimulated echo acquisition mode scans and their comparison. *Quant Imaging Med Surg.* 2021;11:9-20. doi:10.21037/QIMS-20-517
8. Choi C, Ganji S, Hulsey K, et al. A comparative study of short- and long-TE 1 H MRS at 3 T for *in vivo* detection of 2-hydroxyglutarate in brain tumors. *NMR Biomed.* 2013;26:1242-1250. doi:10.1002/nbm.2943
9. Dobberthien BJ, Tessier AG, Yahya A. Improved resolution of glutamate, glutamine and γ -aminobutyric acid with optimized

point-resolved spectroscopy sequence timings for their simultaneous quantification at 9.4 T. *NMR Biomed.* 2018;31:1-12. doi:10.1002/nbm.3851

10. Najac C, Boer VO, Kan HE, Webb AG, Ronen I. Improved detection limits of J-coupled neurometabolites in the human brain at 7 T with a J-refocused sLASER sequence. *NMR Biomed.* 2022;35:1-16. doi:10.1002/nbm.4801
11. Tkáč I, Starčuk Z, Choi IY, Gruetter R. In vivo 1H NMR spectroscopy of rat brain at 1 ms echo time. *Magn Reson Med.* 1999;41:649-656. doi:10.1002/(SICI)1522-2594(199904)41:4<649::AID-MRM2>3.0.CO;2-G
12. Deelchand DK, Berrington A, Noeske R, et al. Across-vendor standardization of semi-LASER for single-voxel MRS at 3T. *NMR Biomed.* 2019;34:1-11. doi:10.1002/nbm.4218
13. Öz G, Tkáč I. Short-echo, single-shot, full-intensity proton magnetic resonance spectroscopy for neurochemical profiling at 4 T: validation in the cerebellum and brainstem. *Magn Reson Med.* 2011;65:901-910. doi:10.1002/mrm.22708
14. Simpson R, Devenyi GA, Jezzard P, Hennessy TJ, Near J. Advanced processing and simulation of MRS data using the FID appliance (FID-A)—an open source, MATLAB-based toolkit. *Magn Reson Med.* 2017;77:23-33. doi:10.1002/mrm.26091
15. Provencher SW. Automatic quantitation of localized in vivo 1H spectra with LCModel. *NMR Biomed.* 2001;14:260-264.
16. Považan M, Hangel G, Strasser B, et al. Mapping of brain macromolecules and their use for spectral processing of 1H-MRSI data with an ultra-short acquisition delay at 7T. *Neuroimage.* 2015;121:126-135. <https://zenodo.org/records/3906754#.XvOWVvJ7n7e> doi:10.1016/j.neuroimage.2015.07.042
17. Edden RAE, Puts NAJ, Harris AD, Barker PB, Evans CJ. Gannet: a batch-processing tool for the quantitative analysis of gamma-aminobutyric acid-edited MR spectroscopy spectra. *J Magn Reson Imaging.* 2014;40:1445-1452. doi:10.1038/jid.2014.371
18. Gasparovic C, Song T, Devier D, et al. Use of tissue water as a concentration reference for proton spectroscopic imaging. *Magn Reson Med.* 2006;55:1219-1226. doi:10.1002/mrm.20901
19. R Core Team. *R: A language and environment for statistical computing.* R Foundation for Statistical Computing; 2020. <https://www.R-project.org/>
20. Bland M. How should I calculate a within-subject coefficient of variation? 2006. <https://www-users.york.ac.uk/~mb55/meas/cv.htm> Accessed June 13, 2024.
21. Deelchand DK, Auerbach EJ, Kobayashi N, Marjańska M. Transverse relaxation time constants of the five major metabolites in human brain measured in vivo using LASER and PRESS at 3 T. *Magn Reson Med.* 2018;79:1260-1265. doi:10.1002/mrm.26826
22. Cudalbu C, Behar KL, Bhattacharyya PK, et al. Contribution of macromolecules to brain 1H MR spectra: experts' consensus recommendations. *NMR Biomed.* 2021;34:1-24. doi:10.1002/nbm.4393

SUPPORTING INFORMATION

Additional supporting information may be found in the online version of the article at the publisher's website.

Data S1. Supporting information.

How to cite this article: Bell TK, Goerzen D, Near J, Harris AD. Examination of methods to separate overlapping metabolites at 7T. *Magn Reson Med.* 2024;1-11. doi: 10.1002/mrm.30293